




Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.



Contents lists available at ScienceDirect

Travel Medicine and Infectious Disease

journal homepage: www.elsevier.com/locate/tmaid

Correspondence

Comparative evaluation of four rapid SARS-CoV-2 antigen detection tests using universal transport medium



ARTICLE INFO

Keywords

SARS-CoV-2
 COVID-19
 Diagnosis
 Rapid diagnostic test
 Antigen detection

Dear Editor,

The containment of the ongoing COVID-19 pandemic requires reliable detection of COVID-19 cases, currently done by real-time reverse-transcription polymerase chain reaction (RT-PCR) [1]. The gap between the number of samples and laboratory capacities to perform RT-PCR in a timely manner, however, is a major limitation of the public health response to COVID-19 [1]. Therefore, there is a critical demand for alternative detection methods, especially rapid diagnostic tests (RDTs), which due to their ease of use might serve as point-of-care tests in community-based settings [2]. Antibody detection tests for SARS-CoV-2 are limited by the delay in humoral immune response, whereas newly developed assays targeting viral antigens have the potential for early diagnosis [2]. However, the accuracy and real-world performance of such assays is unknown and their validation is therefore of high priority [2]. Here we present a head-to-head comparison of four novel antigen-based RDTs for the detection of SARS-CoV-2 from respiratory specimens.

Two of the evaluated assays were based on classical immunochromatography and two used immunofluorescence (Table 1). Samples were obtained from patients presenting respiratory symptoms and/or fever between March 16 and April 26, 2020, and consisted of nasoro-pharyngeal swabs placed in universal transport medium (UTM-RT® System, Copan Diagnostics, Murrieta, USA). UTM specimens were initially examined for SARS-CoV-2 by COVID-19 Genesig® Real-Time PCR (Primerdesign Ltd., Chander's Ford, UK). Exponential amplification curves with cycle threshold (Ct) values ≤ 40 were considered positive. Samples were kept at -80°C before testing by 1) Biocredit COVID-19 Antigen Test (RapiGen Inc.), 2) StrongStep® COVID-19 Antigen Test (Liming Bio-Products Co.), 3) Huaketai New Coronavirus (Savant Biotechnology Co.), and 4) Diagnostic Kit for 2019-nCoV Ag Test (Bioeasy Biotechnology Co.). Noteworthy, the test protocol deviated from manufacturer's instructions by using an equivalent volume of UTM (instead of the recommended test buffer), as previously described [3,4]. Samples were selected by convenience among the 5276 respiratory specimens processed for SARS-CoV-2 during the study period. Due to the shortage of test kits, a 2:1 distribution of positive to negative samples

was chosen. Seventeen positive specimens had been used in a previous evaluation [3].

Assays were tested in parallel from the same sample, performed under BSL2 conditions by the same trained technician, who was blinded to RT-PCR results. Assays with visual output were read by two independent observers, conferring with a third in case of disagreement. RT-PCR served as reference method; for samples with discordant result, tests were repeated. Demographic and clinical data were obtained from mandatory notification forms and analysed anonymously. Samples with high viral loads (Ct value ≤ 25) were compared to those with low viral load (Ct values > 25), as previously described [4]. Statistical analysis considered sensitivity, specificity, accuracy, and Kappa coefficient using standard formulas, and Wilson score Confidence Interval at 95% (OpenEpi version 3.01, GraphPad Prism version 8.4.2). Study materials were purchased with laboratory funds, except for Savant RDT, which was provided free-of-charge through a local provider. The study was approved by the institutional review board (Comité Ético Científico, Facultad de Medicina Clínica Alemana, Universidad del Desarrollo, Santiago, Chile) and need for informed consent was waived.

The study included a total of 111 samples from symptomatic patients; 55% were female, with a median age of 40 years. Eighty specimens were RT-PCR positive, representing 22% of all positives during the study period; 31 samples were RT-PCR negative. The median duration from symptom onset to sampling was 2 days (IQR 1–5 days); 88% of specimens (96/109; missing data, $n = 2$) were taken during the first week of symptoms. Ct values ranged from 10.7 to 37.7 (mean, 22.5). Test performances showed significant differences (Table 1). The evaluation of the Liming Bio kit was stopped after 19 samples, due to its poor results. The other three assays had sensitivities ranging from 16.7% (Savant) to 85% (Bioeasy) and a specificity of 100%. Sensitivities were significantly higher in specimens with high viral loads (Ct values ≤ 25) for RapiGen (84.9%) and Bioeasy (100%) (Table 1). Concordance between these two tests was 82%, while their agreement with Savant was 67% and 50%, respectively. The visual readout of RapiGen was clear, regardless of the intensity of bands. The interpretation of Savant, requiring a UV flashlight provided by the manufacturer, was difficult; its sensitivity might have been higher using an automated reader. Bioeasy

<https://doi.org/10.1016/j.tmaid.2020.101942>

Received 23 June 2020; Received in revised form 23 November 2020; Accepted 24 November 2020

Available online 2 December 2020

1477-8939/© 2020 Elsevier Ltd. All rights reserved.

Table 1
Characteristics and performance of four rapid SARS-CoV-2 antigen-detection tests.

| Characteristics | Test N°1 | Test N°2 | Test N°3 | Test N°4 |
|------------------------------------|---|---|---|---|
| Commercial name | Biocredit One Step SARS-CoV-2 Antigen Test | StrongStep® COVID-19 Antigen Test | Huaketai New Coronavirus (SARS-CoV-2) N Protein Detection Kit (FIA) | Diagnostic Kit for 2019-Novel Coronavirus (2019-nCoV) Ag Test (FIA) |
| Manufacturer | RapiGen Inc., Anyang-si, Gyeonggi-do, Rep. of Korea | Liming Bio-Products Co., Jiangsu, China | Savant Biotechnology Co., Beijing, China | Bioeasy Biotechnology Co., Shenzhen, China |
| Catalogue N° (lot N°) | G61RHA20 (H073001SD) | 500200 (2003014) | BCT-HKT-050 (20031501) | YRLF04401025 (2002N408) |
| Certification ^a | CE-IVD | CE-IVD | CE-IVD | CE-IVD |
| Primary specimen ^b | NP/OP swab | NP/OP swab | Throat swab | NP/OP swab, sputum |
| Incubation (ambient) ^b | 5–8 minutes | 15–20 minutes | 15 minutes ± 1 minute | 10 minutes ± 0 minutes |
| Readout ^b | Visual: coloured bands | Visual: coloured bands | Visual: fluorescent bands ^c | Automated: fluorescence reader |
| Performance^d | | | | |
| Sample size (n) | 109 ^e | 19 ^f | 109 ^e | 111 |
| Sensitivity | 62% (49/79) CI95% 51–71.9 | 0% (0/9) CI95% 0–29.9 | 16.7 (13/78) CI95% 10–16.5 | 85% (68/80) 75.6–91.2 |
| Specificity | 100% (30/30) CI95% 88.7–100 | 90% (9/10) CI95% 59.6–98.2 | 100% (31/31) CI95% 89–100 | 100% (31/31) CI95% 89–100 |
| Accuracy | 72.5% | 47.4% | 40.4% | 89.2% |
| Kappa coefficient | 0.5 | –0.1 | 0.1 | 0.8 |
| Sensitivity, high VL ^g | 84.9% (45/53) CI95% 72.9–92.1 | NA | 21.2% (11/52) CI95% 12.2–34 | 100% (54/54) CI95% 93.4–100 |
| Sensitivity, low VL ^h | 15.4% (4/26) CI95% 17.5–37.7 | NA | 7.7% (2/26) CI95% 2.1–24.1 | 53.8% (14/26) CI95% 25.5–37.4 |
| Mean Ct of false negatives (range) | 29.6 (17.5–37.7) | NA | 21.9 (10.7–37.7) | 34.4 (25.5–37.4) |

FIA, fluorescence immune assay; NP, nasopharyngeal; OP, oropharyngeal; UTM, universal transport medium.

^a According to <https://www.finddx.org/covid-19/pipeline>.

^b According to manufacturer's recommendation.

^c Using UV flashlight recommended and provided by manufacturer.

^d Study protocol included deviation from manufacturer's instructions (see text).

^e Two invalid results were excluded.

^f Testing was suspended after 19 samples due to poor test performance.

^g Samples with high viral loads (Ct ≤ 25).

^h Samples with low viral loads (Ct > 25).

cassettes were interpreted by a desktop instrument with options for QR coding, printing, connectivity to laboratory information systems. Overall, the three systems were easy to use and gave a qualitative result in 10–20 minutes.

Although our study directly compared the assays from the same sample material, the off-label use of UTM might have influenced test results. However, some of the assays showed favourable overall sensitivities, suggesting the potential use of antigen-based RDTs as alternative (or adjunct) tools to RT-PCR. As in other studies [3,5], the performance was significantly higher in specimens with high viral loads (Ct ≤ 25). Since culture studies have shown a significant reduction of infectivity with low viral counts (Ct > 24) [6], antigen testing might play a crucial role within strategies aiming to determine the contagiousness of infected individuals.

In conclusion, the study demonstrated a significant heterogeneity of test performance, which might have been influenced by the use of UTM as a non-validated sample material. The results emphasize that rapid antigen detection has the potential to serve as an alternative diagnostic method, especially as a screening tool for patients with high viral loads during early and infective stages of infection.

Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

CRediT authorship contribution statement

T. Weitzel: Conceptualization, Formal analysis, Methodology, Project administration, Validation, Writing - original draft, Writing - review & editing. **P. Legarraga:** Data curation, Formal analysis, Supervision, Validation, Writing - review & editing. **M. Iruetagoiena:** Data curation, Formal analysis, Validation, Writing - review & editing.

G. Pizarro: Data curation, Investigation, Writing - review & editing. **V. Vollrath:** Supervision, Validation, Writing - review & editing. **R. Araos:** Validation, Writing - review & editing. **J.M. Munita:** Validation, Writing - review & editing. **L. Porte:** Conceptualization, Data curation, Formal analysis, Investigation, Methodology, Project administration, Supervision, Validation, Visualization, Writing - review & editing.

Declaration of competing interest

There is no conflict of interest.

Acknowledgements

We thank Sabine Dittrich (FIND, Geneva) for the critical review of the manuscript.

References

- [1] World Health Organization WHO. Laboratory testing strategy recommendations for COVID-19. Interim guidance. https://apps.who.int/iris/bitstream/handle/10665/331509/WHO-COVID-19-lab_testing-2020.1-eng.pdf. [Accessed 18 November 2020].
- [2] World Health Organization WHO. Advice on the use of point-of-care immunodiagnostic tests for COVID-19. Scientific Brief. <https://www.who.int/news-room/commentaries/detail/advice-on-the-use-of-point-of-care-immunodiagnostic-tests-for-covid-19>. [Accessed 18 November 2020].
- [3] Porte L, Legarraga P, Vollrath V, Aguilera X, Munita JM, Araos R, et al. Evaluation of novel antigen-based rapid detection test for the diagnosis of SARS-CoV-2 in respiratory samples. *Int J Infect Dis* 2020;99:328–33.
- [4] Mertens P, De Vos N, Martiny D, Jassoy C, Mirazimi A, Cuypers L, et al. Development and potential usefulness of the COVID-19 Ag Respi-Strip diagnostic assay in a pandemic context. *Front Med* 2020;7:225.
- [5] Diao B, Wen K, Chen J, Liu Y, Yuan Z, Han C, et al. Diagnosis of acute respiratory syndrome coronavirus 2 infection by detection of nucleocapsid protein [preprint]. <https://www.medrxiv.org/content/10.1101/2020.03.07.20032524v2>.
- [6] Bullard J, Dust K, Funk D, Strong JE, Alexander D, Garnett L, et al. Predicting infectious SARS-CoV-2 from diagnostic samples. *Clin Infect Dis* 2020. <https://doi.org/10.1093/cid/cia638>.

Thomas Weitzel^{*}

Laboratorio Clínico, Clínica Alemana, Facultad de Medicina Clínica
Alemana, Universidad del Desarrollo, Santiago, Chile
Instituto de Ciencias e Innovación en Medicina (ICIM), Facultad de
Medicina Clínica Alemana, Universidad del Desarrollo, Santiago, Chile

Paulette Legarraga¹, Mirentxu Iruretagoyena¹, Gabriel Pizarro,
Valeska Vollrath

Laboratorio Clínico, Clínica Alemana, Facultad de Medicina Clínica
Alemana, Universidad del Desarrollo, Santiago, Chile

Rafael Araos, José M. Munita

Instituto de Ciencias e Innovación en Medicina (ICIM), Facultad de
Medicina Clínica Alemana, Universidad del Desarrollo, Santiago, Chile
Servicio de Infectología, Clínica Alemana de Santiago, Facultad de Medicina
Clínica Alemana, Universidad del Desarrollo, Santiago, Chile

Millennium Initiative for Collaborative Research On Bacterial Resistance
(MICROB-R), Santiago, Chile

Lorena Porte^{**}

Laboratorio Clínico, Clínica Alemana, Facultad de Medicina Clínica
Alemana, Universidad del Desarrollo, Santiago, Chile

* Corresponding author. Laboratorio Clínico, Clínica Alemana, Av.
Vitacura 5951, Santiago, Chile.

** Corresponding author. Laboratorio Clínico, Clínica Alemana, Av.
Vitacura 5951, Santiago, Chile.

E-mail address: thomas.weitzel@gmail.com (T. Weitzel).

E-mail address: lporte@alemana.cl (L. Porte).

¹ These authors contributed equally.